

DEC 09 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130183✓
Name	HUTT, William T.	ID#	3359✓
Date	12/31/10✓		
Agency	Homer Police/Jail	Phone #	235-3150
			235-3151 (fax)
Instrument Location	4060 Heath St Homer, AK 99603		
Alco S/N	X124630✓	Target Value	0.076✓
High Pressure	1300 psi		
Alco Test Values	.075✓	.075✓	
	1 st Alco	2 nd Alco	
Signature	William T. Hutt		
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of DEC, 2010.

Carolyn M. Notland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

OCTOBER 31, 2010

OPERATOR'S NAME:

HUTT, WILLIAM T

OPERATOR'S NUMBER: 3359

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

ETOH

O.L. #: 01234567

DEPT/AGENCY: HOMI

CASE/REPORT: 10-0000

TEST TYPE: U

ALCO TARGET VALUE: 0.076

ALCO S/N: X124630

— BREATH ANALYSIS —

.076 ADJUSTED FOR 28.69 in
ALCO TARGET .072 15:45
BLANK TEST .000 15:46
INTERNAL STANDARD VERIFIED 15:46
ALCO TU 28.69 in .075 15:47
BLANK TEST .000 15:47
SUBJECT SAMPLE .000 15:48
BLANK TEST .000 15:49
ALCO TU 28.69 in .075 15:49
BLANK TEST .000 15:50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

OCTOBER 31, 2010
TIME 16:10

WTHX

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 28.68 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢